



Laerdal
helping save lives


CONFORMITY DECLARATION - MDD

We hereby declare that "**Laerdal Pocket Mask, model 8200xx and 8299xx**", with accessories, manufactured by Laerdal Medical AS, p.o.box 377, N-4002 Stavanger, is manufactured in conformity with the product described in the technical file. The device is in class I (annex VII) according to the classification rules of annex IX.


Compliance is claimed for the essential requirements of Council directive 93/42/EEC regarding the harmonization of the memberstates' regulations regarding medical devices.

The basic product standards used are *ISO 8382:1988 Resuscitators Intended for Use With Humans* and *AS 4259-1995 Ancillary devices for expired air resuscitation*.

For company Laerdal Medical AS



Hilde Tertnes



Date

Laerdal Medical AS
p.o.box 377
N-4002 Stavanger

Telephone: 51511700 Fax: 51511788